

## Overview

### Useful For

Diagnosing insulinoma, when used in conjunction with proinsulin and C-peptide measurements

Management of diabetes mellitus

### Method Name

Electrochemiluminescence Immunoassay (ECLIA)

### NY State Available

Yes

## Specimen

### Specimen Type

Serum

### Ordering Guidance

Patients on insulin therapy may develop anti-insulin antibodies. These antibodies may interfere in the assay system, causing inaccurate results. In such individuals, measurement of free insulin INSFT / Insulin, Free and Total, Serum should be performed.

### Specimen Required

#### Patient Preparation:

1. Patient should be fasting.
2. **For 12 hours before specimen collection do not** take multivitamins or dietary supplements containing biotin (vitamin B7), which is commonly found in hair, skin, and nail supplements and multivitamins.

**Supplies:** Aliquot Tube, 5 mL (T465)

#### Collection Container/Tube:

**Preferred:** Serum gel

**Acceptable:** Red top

**Submission Container/Tube:** Plastic vial

**Specimen Volume:** 1 mL

#### Collection Instructions:

1. Avoid hemolysis
2. Label specimens with corresponding collection times.

3. Centrifuge and aliquot serum into plastic vial within 2 hours of collection.

**Additional Information:** If multiple specimens are drawn, send separate order for each specimen.

### Specimen Minimum Volume

0.75 mL

### Reject Due To

Gross hemolysis	Reject
Gross lipemia	OK
Autopsy specimen	Reject

### Specimen Stability Information

Specimen Type	Temperature	Time	Special Container
Serum	Frozen (preferred)	180 days	
	Refrigerated	7 days	

## Clinical and Interpretive

### Clinical Information

Insulin is a hormone produced by the beta cells of the pancreas. It regulates the uptake and utilization of glucose and is also involved in protein synthesis and triglyceride storage.

Type 1 diabetes (insulin-dependent diabetes) is caused by insulin deficiency due to destruction of insulin-producing pancreatic islet (beta) cells. Type 2 diabetes (noninsulin-dependent diabetes) is characterized by resistance to the action of insulin (insulin resistance).

Insulin levels may be increased in patients with pancreatic beta cell tumors (insulinoma).

### Reference Values

2.6-24.9 mIU/mL

For SI unit Reference Values, see <https://www.mayocliniclabs.com/order-tests/si-unit-conversion.html>

### Interpretation

During prolonged fasting, when the patient's glucose level is reduced to <40 mg/dL, an elevated insulin level plus elevated levels of proinsulin and C-peptide suggest insulinoma.

Insulin levels generally decline in patients with type 1 diabetes mellitus.

In the early stage of type 2 diabetes, insulin levels are either normal or elevated. In the late stage of type 2 diabetes, insulin levels decline.

In normal individuals, insulin levels parallel blood glucose levels.

---

To compare insulin and C-peptide concentrations (ie, insulin to C-peptide ratio):

-Convert insulin to pmol/L: insulin concentration in mIU/mL x 6.945 = insulin concentration in pmol/L.

-Convert C-peptide to pmol/L: C-peptide concentration in ng/mL x 331 = C-peptide concentration in pmol/L.

### Cautions

Human antimouse antibodies (HAMA) may interfere with the assay.

This assay has 100% cross-reactivity with recombinant human insulin (Novolin R and Novolin N). It does not recognize other commonly used analogues of injectable insulin (ie, insulin lispro, insulin aspart, and insulin glargine).

### Clinical Reference

1. Threatte GA, Henry JB: Carbohydrates. In: Henry JB, ed. Clinical Diagnosis and Management by Laboratory Methods. 19th ed. WB Saunders Company; 1996:194-207
2. Sacks DB: Diabetes mellitus. In: Rifai N, Horvath AR, Wittwer CT, eds. Tietz Textbook of Clinical Chemistry and Molecular Diagnostics 6th ed. Elsevier; 2018:1160-1200

### Performance

#### Method Description

This insulin method is a sandwich electrochemiluminescence immunoassay that employs a biotinylated monoclonal insulin-specific antibody and a monoclonal insulin-specific antibody. Insulin in the specimen reacts with both the biotinylated monoclonal insulin-specific antibody (mouse) and the monoclonal insulin-specific antibody (mouse) labeled with a ruthenium complex, forming a sandwich complex. Streptavidin-coated microparticles are added and the mixture is aspirated into the measuring cell where the microparticles are magnetically captured onto the surface of the electrode. Unbound substances are then removed with ProCell. Application of voltage to the electrode induces the chemiluminescent emission, which is then measured. (Package insert: Roche Insulin reagent. Roche Diagnostics; [V1 10/2010](#))

#### PDF Report

No

#### Day(s) Performed

Monday through Saturday

#### Report Available

Same day/1 to 3 days

#### Specimen Retention Time

3 months

#### Performing Laboratory Location

Rochester

### Fees and Codes

#### Fees

- Authorized users can sign in to [Test Prices](#) for detailed fee information.

- 
- Clients without access to Test Prices can contact [Customer Service](#) 24 hours a day, seven days a week.
  - Prospective clients should contact their Regional Manager. For assistance, contact [Customer Service](#).

**Test Classification**

This test has been cleared, approved, or is exempt by the US Food and Drug Administration and is used per manufacturer's instructions. Performance characteristics were verified by Mayo Clinic in a manner consistent with CLIA requirements.

**CPT Code Information**

83525

**LOINC® Information**

Test ID	Test Order Name	Order LOINC Value
INS	Insulin, S	27873-9

Result ID	Test Result Name	Result LOINC Value
INS	Insulin, S	27873-9