

**Overview**

**Useful For**

Supporting the diagnosis of Lyme disease in conjunction with serologic testing

**Testing Algorithm**

See [Acute Tick-Borne Disease Testing Algorithm](#) in Special Instructions

**Special Instructions**

- [Acute Tick-Borne Disease Testing Algorithm](#)

**Method Name**

Real-Time Polymerase Chain Reaction (PCR)/DNA Probe Hybridization

**NY State Available**

Yes

**Specimen**

**Specimen Type**

Whole Blood EDTA

**Ordering Guidance**

This assay does not detect *Borrelia miyamotoi*. If infection with this organism is suspected, order BMIYB / *Borrelia miyamotoi* Detection PCR, Blood or BMIYC / *Borrelia miyamotoi* Detection PCR, Spinal Fluid.

**Specimen Required**

**Container/Tube:** Lavender top (EDTA)

**Specimen Volume:** 1 mL

**Forms**

If not ordering electronically, complete, print, and send a [Microbiology Test Request](#) (T244) with the specimen.

**Specimen Minimum Volume**

0.3 mL

**Reject Due To**

Gross hemolysis	OK
-----------------	----

**Specimen Stability Information**

Specimen Type	Temperature	Time	Special Container
Whole Blood EDTA	Refrigerated (preferred)	7 days	
	Frozen	7 days	

## Clinical and Interpretive

### Clinical Information

Lyme disease is a multisystem and multistage tick-transmitted infection caused by spirochetal bacteria in the *Borrelia burgdorferi* sensu lato (Bbsl) complex.(1) Nearly all human infections are caused by 3 Bbsl species; *B burgdorferi* sensu stricto (hereafter referred to as *B burgdorferi*) is the primary cause of Lyme disease in North America, while *B afzelii* and *B garinii* are the primary causes of Lyme disease in Europe. In 2012, *B mayonii* was identified as a less common cause of Lyme disease in the upper Midwestern United States.(2,3) This organism has only been detected in patients with exposure to ticks in Minnesota and Wisconsin and has not been detected in over 10,000 specimens from patients in other states including regions of the northeast where Lyme disease is endemic.

Lyme disease is the most commonly reported tick-borne infection in Europe and North America, causing an estimated 300,000 cases in the United States each year, and 85,000 cases in Europe.(4,5) The clinical features of Lyme disease are broad and may be confused with various immune and inflammatory disorders. The classic presenting sign of early localized Lyme disease caused by *B burgdorferi* is erythema migrans (EM), which occurs in approximately 80% of individuals. Other early signs and symptoms include malaise, headache, fever, lymphadenopathy, and myalgia. Arthritis, neurological disease, and cardiac disease may be later stage manifestations. Erythema migrans has also been seen in patients with *B mayonii* infection, but diffuse rashes are more commonly reported.(2) The chronic skin condition, acrodermatitis chronica atrophicans, is also associated with *B afzelii* infection.

The presence of EM in the appropriate clinical setting is considered diagnostic for Lyme disease and no confirmatory laboratory testing is needed. In the absence of a characteristic EM lesion, serologic testing is the diagnostic method of choice for Lyme disease.(6) However, serology may not be positive until 1 to 2 weeks after onset of symptoms, and may show decreased sensitivity for detection of infection with *B mayonii*. Therefore, detection of Bbsl DNA using PCR may be a useful adjunct to serologic testing for detection of acute disease. PCR has shown utility for detection of *Borrelia* DNA from skin biopsies of Lyme-associated rashes, and can also be used to detect *Borrelia* DNA from synovial fluid and synovium biopsies. Less commonly, *Borrelia* DNA can be detected in cerebrospinal fluid and blood.(7) In general, blood is not the preferred source for detection of Bbsl DNA by PCR, although it may have increased utility for detection of *B mayonii*, due to the higher levels of observed peripheral spirochetemia with this organism.(2) Lyme PCR should always be performed in conjunction with FDA-approved serologic tests, and results should be correlated with serologic and epidemiologic data and clinical presentation of the patient.(8) The Mayo Clinic Lyme PCR test detects and differentiates the main causes of Lyme disease in North America (*B burgdorferi* and *B mayonii*) and Europe (*B afzelii* and *B garinii*).(2,7)

### Reference Values

Negative

### Interpretation

A positive result indicates the presence of DNA from *Borrelia burgdorferi*, *B mayonii*, *B afzelii*, or *B garinii*, the main agents of Lyme disease.

A negative result indicates the absence of detectable target DNA in the specimen. Due to the diagnostic sensitivity limitations of the PCR assay, a negative result does not preclude the presence of the organism or active Lyme disease.

### Cautions

Serologic tests are recommended for diagnosis of Lyme disease. PCR may play an adjunctive role, but may not detect *Borrelia burgdorferi* DNA from blood in cases of active or chronic disease. The presence of inhibitory substances may also cause a false-negative result. PCR test results should be used as an aid in diagnosis and not

considered diagnostic by themselves. These results should be correlated with serologic and epidemiologic data and clinical presentation of the patient.

Concurrent infections with multiple tick-borne pathogens, including *Ehrlichia muris eauclairensis*, *Anaplasma phagocytophilum*, *Babesia microti*, and *B. miyamotoi* (a relapsing-fever *Borrelia*) have been reported in the United States, and consideration should be given to testing for other pathogens if clinically indicated.

Â

This assay detects most members of the *B burgdorferi* sensu lato complex, including *B. andersonii*, *B. americana*, and *B. bissetii*, which have been rarely detected in humans. Detection of DNA from these organisms would be reported as an atypical result and prompt additional laboratory testing to further identify the DNA present. The sensitivity of this assay for detecting these organisms has not been determined.

This assay also detects some members of the *B burgdorferi* sensu lato (Bbsl) complex that are not considered to be human pathogens, but may be found in ticks and other animals. Therefore, this assay should not be used to test nonhuman specimens.

### Supportive Data

The following validation data supports the use of this assay for clinical testing.

Analytical Sensitivity/Limit of Detection (LoD):

The lower LoD is approximately 300 to 1,000 genomic copies/mL in cerebrospinal fluid, tissue, blood, and synovial fluid.

Accuracy/Diagnostic Sensitivity and Specificity:

Spiking studies of whole organism in whole blood (spiked near the approximate limit of detection) showed 100% recovery.

Analytical Specificity:

No PCR signal was obtained from the extracts of 22 bacterial, viral, parasitic, and fungal isolates that can cause symptoms similar to Lyme disease including; *Rickettsia rickettsii*, *R. typhi*, *Ehrlichia canis*, *Babesia microti*, *Plasmodium falciparum*, *P. vivax*, *Bartonella henselae*, *Bartonella quintana*, Herpes simplex virus, and *Toxoplasma gondii*. Relapsing fever borreliae (including *Borrelia miyamotoi*) are also not detected with this assay.

Precision:

Interassay precision was 100% and intra-assay precision was 100%.

Reference Range:

The reference range for this assay is negative. This assay is only to be used for patients with a clinical history and symptoms consistent with Lyme, and must be interpreted in the context of serologic tests, which are the gold standard for diagnosis of Lyme disease. This test should not be used to screen asymptomatic patients.

Reportable Range:

This is a qualitative assay, and the results are reported as negative or positive for targeted *Borrelia burgdorferi*.

---

## Clinical Reference

1. Stanek G, Wormser GP, Gray J, Strle F: Lyme borreliosis. *Lancet* 2012;379(9814):461-473
2. Pritt BS, Mead PS, Johnson, DK, et al: Identification of a novel pathogenic *Borrelia* species causing Lyme borreliosis with unusually high levels of spirochetemia: a descriptive study. *Lancet Infect Dis* 2016 May;16(5):556-564
3. Pritt BS, Respcio-Kingry LB, Sloan LM, et al: *Borrelia mayonii* sp. nov., a member of the *Borrelia burgdorferi* sensu lato complex, detected in patients and ticks in the upper midwestern United States. *Int J Sys Evol Microbiol* 2016;66(11):4878-4880
4. Hinckley AF, Connally NP, Meek JI, et al: Lyme disease testing by large commercial laboratories in the United States. *Clin Infect Dis* 2014;59(5):676-681
5. Lindgren E, Jaenson TGT: Lyme borreliosis in Europe: influences of climate and climate change, epidemiology, ecology and adaptation measures. Copenhagen, Denmark: World Health Organization; 2006
6. Centers for Disease Control and Prevention. Recommendations for test performance and interpretation from the Second National Conference on Serologic Diagnosis of Lyme Disease. *MMWR Morb Mortal Wkly Rep* 1995;44(31):590-591
7. Babady NE, Sloan LM, Vetter EA, et al: Percent positive rate of Lyme real-time polymerase chain reaction in blood, cerebrospinal fluid, synovial fluid, and tissue. *Diagn Microbiol Infect Dis* 2008;62(4):464-466
8. CDC: Recommendation for test performance and interpretation. From second national conference on serological diagnosis of lyme disease. *MMWR Morb Mortal Wkly Rep* 1996;45:481-484

## Performance

### Method Description

Nucleic acid is extracted from clinical specimens using the automated MagNA Pure LC instrument system. The extract is then transferred wells of a 96-well plate for amplification. The LightCycler is an automated instrument that amplifies and monitors the development of target nucleic acid (amplicon) after each cycle of PCR. The DNA target for PCR assay is the 283-bp plasminogen-binding protein gene (*OppA2*), which is present at a frequency of 1 copy per organism in all 4 confirmed pathogenic species of the *Borrelia burgdorferi* sensu lato genogroup (*B burgdorferi* sensu stricto, *B afzelii*, *B garinii*, and *B mayonii*). A specific base pair DNA target sequence is amplified by PCR. The detection of amplicon is based on fluorescence resonance energy transfer (FRET), which utilizes 1 hybridization probe with a donor fluorophore, fluorescein, at the 3' end, and a second hybridization probe with an acceptor fluorophore, LC-Red 610, at the 5' end. When the target amplicon is present, the LC-Red 610 emits a measurable and quantifiable light signal at a specific wavelength. Presence of the specific organism nucleic acid may be confirmed by performing a melting curve analysis of the amplicon. Using features of the melting curve analysis, the assay primers and specific hybridization probes are able to detect and differentiate *B burgdorferi* sensu stricto from *B mayonii*, *B afzelii*, and *B garinii*, although the melting curve analysis cannot differentiate between *B afzelii* and *B garinii*. Each assay run can be completed within 60 minutes.(Cockerill FR, Uhl FR: Applications and challenges of real-time PCR for the clinical microbiology laboratory. In *Rapid Cycle Real-Time PCR*. Edited by U Reischl, C Wittwer, F Cockerill. Springer, NY 2002)

### PDF Report

No

**Day(s) Performed**

June through November: Monday through Saturday

December through May: Monday through Friday

**Report Available**

1 to 4 days

**Specimen Retention Time**

1 week

**Performing Laboratory Location**

Rochester

**Fees and Codes**
**Fees**

- Authorized users can sign in to [Test Prices](#) for detailed fee information.
- Clients without access to Test Prices can contact [Customer Service](#) 24 hours a day, seven days a week.
- Prospective clients should contact their Regional Manager. For assistance, contact [Customer Service](#).

**Test Classification**

This test was developed, and its performance characteristics determined by Mayo Clinic in a manner consistent with CLIA requirements. This test has not been cleared or approved by the US Food and Drug Administration.

**CPT Code Information**

87476

87798 x 2

87999 (if appropriate for government payers)

**LOINC® Information**

Test ID	Test Order Name	Order LOINC Value
PBORB	Lyme Disease PCR, B	90892-1

Result ID	Test Result Name	Result LOINC Value
56080	B. burgdorferi PCR, B	94247-4
38290	B. mayonii PCR, B	94248-2
38291	B. garinii/B. afzelii PCR, B	94249-0
38340	Lyme Disease PCR Comment	59464-8