

Test Definition: FDFEN

Dog Fennel (Anthemis cotula) IgE

Overview

Method Name

ImmunoCAP FEIA

NY State Available

Yes

Specimen

Specimen Type

Serum

Specimen Required

Specimen Type: Serum Container/Tube: Red or SST Specimen Volume: 0.5 mL

Collection Instructions: Draw blood in a plain red-top tube(s), serum gel tube is acceptable. Spin down and send 0.5 mL of serum refrigerated in a plastic vial.

Specimen Minimum Volume

0.5 mL

Reject Due To

Gross	Reject
hemolysis:	
Thawing:	Warm OK; Cold OK
Gross lipemia:	Reject

Specimen Stability Information

Specimen Type	Temperature	Time	Special Container
Serum	Refrigerated (preferred)	28 days	
	Ambient	28 days	
	Frozen	365 days	

Clinical & Interpretive



Test Definition: FDFEN

Dog Fennel (Anthemis cotula) IgE

Reference Values

<0.35 kU/L

Interpretation

		-
Class	IgE (kU/L)	Comment
0	<0.10	Negative
0/1	0.10-0.34	Equivocal/Borderline
1	0.35-0.69	Low Positive
2	0.70-3.49	Moderate Positive
3	3.50-17.49	High Positive
4	17.50-49.99	Very High Positive
5	50.00-99.99	Very High Positive
6	>99.99	Very High Positive

Performance

PDF Report

No

Day(s) Performed

Monday through Friday

Report Available

3 to 6 days

Performing Laboratory Location

Eurofins Viracor

Fees & Codes

Fees

- Authorized users can sign in to <u>Test Prices</u> for detailed fee information.
- Clients without access to Test Prices can contact <u>Customer Service</u> 24 hours a day, seven days a week.
- Prospective clients should contact their account representative. For assistance, contact <u>Customer Service</u>.

Test Classification

This test was developed and its performance characteristics determined by Eurofins Viracor. It has not been cleared or approved by the U.S. Food and Drug Administration.

CPT Code Information



Test Definition: FDFEN

Dog Fennel (Anthemis cotula) IgE

86003

LOINC® Information

Test ID	Test Order Name	Order LOINC® Value
FDFEN	Dog Fennel (Anthemis cotula) IgE	7287-6

Result ID	Test Result Name	Result LOINC® Value
Z5606	Dog Fennel (Anthemis cotula) IgE	7287-6
Z5607	Class	Unable to Verify