

## Overview

### Useful For

Detection or monitoring of IgA monoclonal gammopathies and IgA-related immune deficiencies

### Testing Algorithm

The following algorithms are available:

- [Celiac Disease Comprehensive Cascade Test Algorithm](#)
- [Celiac Disease Diagnostic Testing Algorithm](#)
- [Celiac Disease Gluten-Free Cascade Test Algorithm](#)
- [Celiac Disease Routine Treatment Monitoring Algorithm](#)
- [Celiac Disease Serology Cascade Test Algorithm](#)

### Special Instructions

- [Celiac Disease Diagnostic Testing Algorithm](#)
- [Celiac Disease Comprehensive Cascade Test Algorithm](#)
- [Celiac Disease Gluten-Free Cascade Test Algorithm](#)
- [Celiac Disease Routine Treatment Monitoring Algorithm](#)
- [Celiac Disease Serology Cascade Test Algorithm](#)

### Method Name

Nephelometry

### NY State Available

Yes

## Specimen

### Specimen Type

Serum

### Ordering Guidance

Cascade testing is recommended for celiac disease. Cascade testing ensures that testing proceeds in an algorithmic fashion. The following cascades are available; select the appropriate one for your specific patient situation.

- CDCOM / Celiac Disease Comprehensive Cascade, Serum and Whole Blood: Complete testing including HLA DQ
  - CDSP / Celiac Disease Serology Cascade, Serum: Complete serology testing excluding HLA DQ
  - CDGF / Celiac Disease Gluten-Free Cascade, Serum and Whole Blood: For patients already adhering to a gluten-free diet
- To order individual tests, see [Celiac Disease Diagnostic Testing Algorithm](#).

### Specimen Required

**Supplies:** Sarstedt Aliquot Tube, 5 mL (T914)

**Collection Container/Tube:**  
**Preferred:** Serum gel  
**Acceptable:** Red top  
**Submission Container/Tube:** Plastic vial  
**Specimen Volume:** 1 mL  
**Collection Instructions:** Centrifuge and aliquot serum into a plastic vial.

**Forms**  
If not ordering electronically, complete, print, and send a [Gastroenterology and Hepatology Test Request](#) (T728) with the specimen.

**Specimen Minimum Volume**  
0.5 mL

**Reject Due To**

Gross hemolysis	OK
Gross lipemia	Reject
Gross icterus	OK

**Specimen Stability Information**

Specimen Type	Temperature	Time	Special Container
Serum	Refrigerated (preferred)	28 days	
	Ambient	14 days	
	Frozen	28 days	

**Clinical & Interpretive**

**Clinical Information**

Immunoglobulins are produced by plasma cells as a humoral immune response to contact of the immune system by antigens. The primary reaction after the initial contact is the formation of antibodies of the IgM class, followed later by IgG and IgA antibodies. Quantitative determination of the immunoglobulins can provide important information on the humoral immune status. Decreased serum immunoglobulin concentrations occur in primary immunodeficiency conditions as well as in secondary immune insufficiencies (eg, in advanced malignant tumors, lymphatic leukemia, multiple myeloma, and Waldenstrom disease).

Monoclonal immunoglobulin proliferations in the serum are found in plasmacytomas, Waldenstrom disease, and heavy-chain disease. Monoclonal immunoglobulinemia requires detailed differential diagnostic investigations in addition to the quantitative determination. Local immune reactions result in elevated immunoglobulin levels, particularly IgG, in the cerebrospinal fluid. IgA increases with asparaginase treatment, during pregnancy, with exercise, and in people with alcohol use disorder. It falls with prolonged exposure to benzene and after 1 year's abstinence from drinking alcohol. Diphenylhydantoin, dextran, methyl prednisolone, toluene, xylol, and oral contraceptives may also lower IgA levels. IgM

---

may rise in people with narcotic addiction and after various drug use, as with IgA and IgG.

The gamma globulin band as seen in conventional serum protein electrophoresis consists of 5 immunoglobulins. In normal serum, about 15% is IgA.

Monoclonal gammopathies of all types may lead to a spike in the gamma globulin zone seen on serum protein electrophoresis.

Monoclonal elevations of IgA characterize multiple myeloma.

Decreased immunoglobulin levels are found in patients with congenital deficiencies.

**Reference Values**

0-<5 months: 7-37 mg/dL

5-<9 months: 16-50 mg/dL

9-<15 months: 27-66 mg/dL

15-<24 months: 36-79 mg/dL

2-<4 years: 27-246 mg/dL

4-<7 years: 29-256 mg/dL

7-<10 years: 34-274 mg/dL

10-<13 years: 42-295 mg/dL

13-<16 years: 52-319 mg/dL

16-<18 years: 60-337 mg/dL

> or =18 years: 61-356 mg/dL

**Interpretation**

Increased serum immunoglobulin concentrations occur due to polyclonal or oligoclonal immunoglobulin proliferation in hepatic disease (hepatitis, liver cirrhosis), connective tissue diseases, acute and chronic infections, as well as in the cord blood of neonates with intrauterine and perinatal infections.

Elevation of IgA may occur in monoclonal gammopathies such as multiple myeloma, primary systemic amyloidosis, monoclonal gammopathy of undetermined significance, and related disorders.

Decreased levels are found in patients with primary or secondary immune deficiencies.

**Cautions**

Electrophoresis is usually required to interpret an elevated immunoglobulin class as polyclonal versus monoclonal. Immunofixation is usually required to characterize a monoclonal protein.

If there is a discrete M-peak, the monoclonal protein can be monitored with quantitative immunoglobulins.

If immunoglobulin quantitation is used to monitor the size of a monoclonal protein that is contained in a background of polyclonal immunoglobulins, changes in the immunoglobulin quantitation may reflect changes in the background immunoglobulins; therefore, serum protein electrophoresis should be used to monitor the monoclonal protein.

Results determined by assays using different manufacturers or methods may not be comparable.

Quantitation of specific proteins by nephelometric means may not be possible in lipemic sera due to the extreme light scattering properties of the specimen. Turbidity and particles in the specimen may result in extraneous light scattering signals, resulting in variable specimen analysis.

**Clinical Reference**

1. Webster ADB. Laboratory investigation of primary deficiency of the lymphoid system. In: Clinics in Immunology and Allergy. Vol 5. 3rd ed. WB Saunders Company; 1985:447-468
2. Pinching AJ. Laboratory investigation of secondary immunodeficiency. In: Clinics in Immunology and Allergy. Vol 5. 3rd ed. WB Saunders Company; 1985:469-490
3. Dispenzieri A, Gertz MA, Kyle RA. Distribution of diseases associated with moderate polyclonal gammopathy in patients seen at Mayo Clinic during 1991. Blood. 1997;90:353A
4. Kyle RA, Greipp PR. The laboratory investigation of monoclonal gammopathies. Mayo Clin Proc. 1978;53(11):719-739
5. Ballow M, O'Neil KM. Approach to the patient with recurrent infections. In: Middleton Jr E, Reed CE, Ellis, et al. Allergy: Principles and Practice. Vol 2. 4th ed. Mosby-Year Book, Inc; 1993:1027-1058
6. Kyle RA. Detection of quantitation of monoclonal proteins. Clin Immunol Newsletter. 1990;10(6):84-86
7. Dietzen DJ, Willrich MAV. Amino acids, peptides, and proteins. In: Rifai N, Chiu RWK, Young I, Burnham CAD, Wittwer CT, eds. Tietz Textbook of Laboratory Medicine. 7th ed. Elsevier; 2023:chap 31

**Performance**

**Method Description**

In this Siemens Nephelometer II method, the light scattered onto the antigen-antibody complexes is measured. The intensity of the measured scattered light is proportional to the amount of antigen-antibody complexes in the sample under certain conditions. If the antibody volume is kept constant, the signal behaves proportionally to the antigen volume.

A reference curve is generated by a standard with a known antigen content on which the scattered light signals of the samples can be evaluated and calculated as an antigen concentration. Antigen-antibody complexes are formed when a sample containing antigen and the corresponding antiserum are put into a cuvette. A light beam is generated with a light emitting diode, which is transmitted through the cuvette. The light is scattered onto the immuno-complexes that are present. Antigen and antibody are mixed in the initial measurement, but no complex is formed yet. An antigen-antibody complex is formed in the final measurement.

The result is calculated by subtracting the value of the final measurement from the initial measurement. The distribution of intensity of the scattered light depends on the ratio of the particle size of the antigen-antibody complexes to the radiated wavelength.(Instruction manual: Siemens Nephelometer II, Siemens, Inc.; Version 2.4, 07/2019)

**PDF Report**

No

**Day(s) Performed**

Monday through Friday

Report Available

1 to 3 days

Specimen Retention Time

14 days

Performing Laboratory Location

Mayo Clinic Laboratories - Rochester Superior Drive

Fees & Codes

Fees

- Authorized users can sign in to [Test Prices](#) for detailed fee information.
- Clients without access to Test Prices can contact [Customer Service](#) 24 hours a day, seven days a week.
- Prospective clients should contact their account representative. For assistance, contact [Customer Service](#).

Test Classification

This test has been cleared, approved, or is exempt by the US Food and Drug Administration and is used per manufacturer's instructions. Performance characteristics were verified by Mayo Clinic in a manner consistent with CLIA requirements.

CPT Code Information

82784

LOINC® Information

Test ID	Test Order Name	Order LOINC® Value
IGA	Immunoglobulin A (IgA), S	2458-8

Result ID	Test Result Name	Result LOINC® Value
IGA	Immunoglobulin A (IgA), S	2458-8